




AIM Biotech Luer Connectors

Product Information Sheet
06/30/16 Rev. 1.0

AIM Biotech Luer Connectors are sterile, single-use connectors that bridge the AIM chips to luer fittings such as syringes and tubing adaptors thus enabling modular expansion in the AIM chips. This product comes with AIM Inlet Seals that can prevent the leakage of media from gel inlets when flow is applied in the AIM chips.




PROTOCOL OVERVIEW


Connector Assembly

TIMING 10 min

1. Perform the following steps in a laminar flow hood.
2. Align the nub of the connectors to the corresponding indent at the ports (of AIM chips).



3. Push down the connectors and make sure the contact areas between the connectors and the AIM chips are flush.



Collagen Gel Preparation and Filling

TIMING 50 min

4. Prepare collagen gel according to the pre-determined collagen gel preparation recipe. You may use any hydrogel relevant to your specific application.
5. Fill 10 μ l of collagen solution from either one of the gel inlets and stop near the end of posts. Fill from the other gel inlet until the gel fronts merge.
6. Seal the gel inlets immediately with AIM Inlet Seals by using tweezers.
7. Place the gel-filled chips into a 37 °C incubator and incubate for 30 min.

Media Channel Hydration & Coating


TIMING 70 min

8. After incubation, gently inject 15 μ l of coating solution into the channel that requires coating. Repeat this step for the other channel. Use culture medium instead if coating is not required.
9. Incubate for 1 h or for the optimized time of other coating solutions in a 37 °C incubator.
10. Add 100 μ l of medium into one connector to flush out the coating solution from the opposite connector.

Cell Seeding

TIMING 40 min

11. Prepare cell suspensions with densities ranging from 0.5 to 3 M cells/ml, depending on the application.
12. Remove media from all connectors by using a micropipette. Do not apply negative pressure directly at the connector inlet to prevent the accidental removal of medium from the media channel.



The media levels should be the same as or slightly above the level of connector inlet

13. Add 20 μ l endothelial cell suspension into one connector. Wait for 2 min and then repeat the same procedure for the opposite connector that is attached to the same media channel.
14. Visual inspection under a microscope is recommended. If the cell distribution is not optimal for your application, adjust the density of the cell suspension and repeat the cell seeding step.
15. If another cell type B is to be seeded in the opposite media channel, incubate the chip for at least 30 min after seeding cell type A to allow proper cell attachment on substrates before repeating the cell seeding steps for cell type B.
16. Keep the chips in an incubator.

Media Changing

TIMING 10 min

17. Change media 2 to 4 h after cell seeding (or longer for less adhesive cell types) when the cells have adhered to the substrates.
18. Remove media from all connectors. Add 100 μ l of fresh medium into one connector to flush out the old medium from the opposite connector. Repeat this for the other channel.

Application of Flow

19. **Option 1:** Attach auxiliary reservoirs to the connectors to generate pressure differences between media channels. We recommend using 1ml luer slip syringe barrels as auxiliary reservoirs.
20. **Option 2:** Attach auxiliary luer adaptors and tubing that are connected to syringe pumps to the connectors to control the pressure difference between/within media channels more precisely. We recommend using male luer lock adaptors with 1/16" ID low profile barb and biocompatible tubing. Tubing with sizes ranging from 1/16" (for incompressible tubing) to 1/32" ID (for compressible tubing) may be used. The corresponding needle sizes are 15 and 18 gauge respectively and we recommend using blunt end needles.

Please refer to www.aimbiotech.com/protocols for a detailed explanation of the connector-related and application-specific protocols.

PRECAUTIONS AND WARNINGS

- Luer Connectors are for single use only. Do not reuse. Do not re-sterilize. May cause leakage if reused or re-sterilized.
- Always use aseptic technique while handling the connectors. Do not use if packaging is damaged.
- Not compatible with aromatic hydrocarbons (such as benzene).

SHIPPING AND STORAGE

Luer Connectors are shipped together with AIM Inlet Seals at room temperature. Store in a dry environment at room temperature.

STERILIZATION

Luer Connectors are supplied in a clear doubly-sealed package and sterilized by gamma beam irradiation. The connectors are sterile if the packaging is not damaged.

END USER LICENSE AGREEMENT

"Product(s)" means the item(s) acquired hereunder. Products may be Seller-branded accessories, peripherals, consumables or reagents. "Purchaser" means the person or entity acquiring the Product hereunder. "Seller" means AIM Biotech Pte. Ltd., the entity selling the Product hereunder. The Selling entity is identified on the quotation, order acknowledgment or similar communication, or Seller website if the order is being placed electronically at Seller's website. AIM Biotech Pte. Ltd. ("AIM Biotech") owns pending US patent applications and corresponding foreign patents and patent applications covering the Products. The Products are made available to scientists for internal research purposes only, under the terms described below. Use of the Products by commercial entities for purposes other than internal research requires the user to obtain a commercial license as detailed below. The Products are for research use only and may not be used in clinical diagnostics nor in the provision of research services without a separate license. AIM Biotech grants you, the Purchaser, a non-exclusive license to use the Products for research purposes only. The Products are being transferred to the Purchaser in furtherance of, and reliance on, such license. The Purchaser may not use the Products for commercial purposes other than research without a license for such purpose from AIM Biotech. Commercial purposes include: any use of the Products in the manufacture of a commercial product, or in the provision of research, discovery or screening services; any sale or transfer for consideration of the Products; any clinical diagnostic or therapeutic application employing the Products. Access to the Products must be limited solely to those officers, employees and students of the Purchaser's institution who need access to perform the above-described internal research. The Purchaser must inform each such officer, employee and student of the provisions of this license agreement and require them to agree to be bound by the provisions of this license agreement. The Purchaser may not distribute the Products, even to those within your own institution. The Purchaser may not assign, sub-license, rent, lease or otherwise transfer this license agreement or any of the rights or obligation there under, except as expressly permitted by AIM Biotech. This license agreement is effective until terminated. The Purchaser may terminate it at any time by destroying all the Products in the Purchaser's control. It will also terminate automatically if the Purchaser fails to comply with the terms and conditions of this license agreement. The Purchaser shall, upon termination of the license agreement, destroy all Products in its control. For information on commercial use, please contact: The Licensing Coordinator, AIM Biotech Pte. Ltd., 51 Ayer Rajah Crescent #02-12/14, Singapore 139948. Tel/Fax: (+65) 6777-0246, Email: info@aimbiotech.com.

ORDERING INFORMATION

Luer Connectors (Cat. No. AIMLUC) are sold by the pack of 36. Compatible items including chips (Cat. No AIMDAX) and holders (Cat. No AIMHOL) are sold separately.