



PREOMICS

iST Sample Preparation Kit 8x

Urine

Introduction

Sample preparation is one of the essential steps of bottom-up proteomics. The PreOmics iST sample preparation kit is designed to assist researchers achieving best results with few sample preparation steps and little hands-on time. For sample-specific protocols and optimization visit www.preomics.com/downloads or contact info@preomics.com.

Kit Contents

The kit contains everything to perform a complete sample preparation. It includes all chemicals to denature, reduce and alkylate proteins, as well as the enzymes to perform a tryptic digestion and a final peptide cleanup.


Component	Cap	Quantity	Buffer Properties				Description	Storage
			Organic	Acidic	Basic	Volatile		
DIGEST	●	2x					Trypsin/LysC mix to digest proteins.	-20°C
RESUSPEND	●	1x 2 mL				●	Reconstitutes lyophilized proteolytic enzymes.	RT
LYSE	●	1x 1 mL			●		Denatures, reduces and alkylates proteins.	RT
STOP	●	1x 1 mL	●	●		●	Stops the enzymatic activity.	RT
WASH 0	●	1x 2 mL	●	●		●	Cleans peptides from tetrapyrrole contaminants.	RT
WASH 1	●	1x 2 mL	●	●		●	Cleans peptides from hydrophobic contaminants.	RT
WASH 2	●	1x 2 mL		●		●	Cleans peptides from hydrophilic contaminants.	RT
ELUTE	●	1x 2 mL	●		●	●	Elutes the peptides from the cartridge.	RT
LC-LOAD	○	1x 1 mL		●		●	Loads peptides on reversed-phase LC-MS column.	RT
CARTRIDGE		8x					Cartridge for 1 to 100 µg protein starting material.	RT
WASTE		8x					2.0 mL tube for collecting waste after washing steps.	RT
COLLECTION		8x					1.5 mL tube for collecting peptides after elution.	RT
ADAPTER		8x					Enables a cartridge to be placed into a tube.	RT

Pre-Requisites

Common lab equipment is required for the sample preparation.

Equipment	Quantity and Description
PIPETTE	Careful sample handling and pipetting reduces contaminations and improves quantification.
SAMPLE	100 mL urine.
CENTRIFUGAL FILTER	3kDa protein concentration ultrafiltration system (e.g. Merck Millipore ACS500302).
HEATING BLOCK	Two heating blocks are recommended to support protein denaturation and digestion.
CENTRIFUGE	1.5/2.0 mL reaction tube centrifuges are required for loading, washing and elution.
SONICATOR	If the sample contains DNA, shear it by sonication (e.g. Diagenode Bioruptor®).
VACUUM EVAPORATOR	Vacuum manifolds evaporate volatile buffers from the eluate before LC-MS.
ULTRASONIC BATH	Optional: can be used to resuspend peptides.

Procedure



Method

1 LYSE

- 1.1. Load 10 mL urine onto a centrifugal filter (not provided) and concentrate to 1 mL. Repeat until 100 mL are loaded.
- 1.2. Concentrate the sample to 100 µL and add 900 µL LYSE ● to your sample. Mix gently.
- 1.3. Concentrate the solution to 50 µL, mix thoroughly and transfer the sample to a fresh tube.
- 1.4. Place the sample in a HEATING BLOCK (95°C; 1,000 rpm; 10 min). *NOTE 1*
- 1.5. Optional: Spin down droplets (RT; max. 300 rcf; 10 sec).
- 1.6. Use ADAPTER to place CARTRIDGE in WASTE tube. Label all tubes.
- 1.7. Transfer sample to CARTRIDGE and cool down (RT). Be careful not to damage the bottom layer of CARTRIDGE.

2. DIGEST

- 2.1. Add 210 µL RESUSPEND ● to DIGEST ● (1 tube for 4 reactions), shake (RT; 500 rpm; 10 min), pipette up/down.
- 2.2. Add 50 µL DIGEST ● to CARTRIDGE and place it in a pre-heated HEATING BLOCK (37°C; 500 rpm; 1-3 hours).
- 2.3. Add 100 µL STOP ● to CARTRIDGE (precipitation may occur), shake (RT; 500 rpm; 1 min / pipette up/down). *SP*

3. PURIFY

- 3.1. Spin CARTRIDGE in a CENTRIFUGE (3,800 rcf; 1-3 min). If needed, adjust time to ensure complete flow-through.
- 3.2. Add 200 µL WASH 0 ● to CARTRIDGE, repeat step 3.1.
- 3.3. Add 200 µL WASH 1 ● to CARTRIDGE, repeat step 3.1.
- 3.4. Add 200 µL WASH 2 ● to CARTRIDGE, repeat step 3.1. *SP*
- 3.5. Use ADAPTER to place CARTRIDGE in a fresh COLLECTION tube. Label all tubes.
- 3.6. Add 100 µL ELUTE ● to CARTRIDGE, repeat step 3.5., keep flow-through in COLLECTION tube.
- 3.7. Repeat step 3.5., keep flow-through in the same COLLECTION tube.
- 3.8. Discard CARTRIDGE and place COLLECTION tube in a vacuum evaporator (45°C; until completely dry).
- 3.9. Add LC-LOAD ○ to COLLECTION tube. Aim for 1 g/L concentration (e.g. 100 µL to 100 µg protein starting material).
- 3.10. Sonicate COLLECTION tube in an ULTRASONIC BATH (5 min) or shake (RT; 500 rpm; 5 min). *SP*

NOTE1 Volumes of buffers can be adjusted according to protein starting amounts.

Lysis temperature should be between 60-95°C.

Visit our FAQ website for more information: www.preomics.com/faq.

SP - Storage Point:

At this point, close the peptide containing tube or CARTRIDGE using silicon lid.

Peptides can be frozen at -20°C. Storage of peptides should not exceed two weeks at -20°C.

For extended storage, finish the protocol and store at -80°C.

Data analysis

Consider the following as fixed modifications in your database search:

MODIFICATION	DESCRIPTION	COMPOSITION	SPECIFICITY	MASS	UNIMOD #
ALKYLATION	Carbamidomethyl on cysteine	C ₂ H ₃ NO	[C]	+57Da	4

Please refer to www.preomics.com for our General Terms and Conditions.