



## Summary of Mask Testing Data: ASTM F2100-11

Intertek

3933 US Route 11

Cortland, New York 13045

February 12, 2021

Bacterial Filtration Efficacy:

%BFE, all results  $\geq$  99.37% barrier.

**[PASS ALL LEVELS]**

Breathability / Pressure Drop:

Average Delta P, 4.6 mm H<sub>2</sub>O/cm<sup>2</sup>

**[PASS LEVEL 2]**

Face Mask Fluid Barrier Testing:

160mmHg, 32 of 32 Pass

**[PASS LEVEL 3 / ALL LEVELS]**

Flammability:

Class 1, Normal Flammability

**[PASS ALL LEVELS]**

Bioburden:

Could not detect any organisms  
(aerobic, fungal, or total bioburden)

**[PASS]**

Nelson Laboratories, LLC

6280 S. Redwood Rd.

Salt Lake City, UT 84123 U.S.A.


March 25, 2021

Particle Filtration Efficiency

%PFE = 99.7%

**[PASS ALL LEVELS]**

## Summary of Standard



### ASTM F2100-11: Standard Specification for Performance of Materials Used in Medical Face Masks

TEST	LEVEL 1 BARRIER	LEVEL 2 BARRIER	LEVEL 3 BARRIER
ASTM F1862 (Fluid Resistance)	80 mmHg	120 mmHg	160 mmHg
MIL-M-36954 C: $\Delta P$ (Breathability)	< 4 mm H <sub>2</sub> O	< 5 mm H <sub>2</sub> O	< 5 mm H <sub>2</sub> O
ASTM F2101: BFE (Filtration 3 $\mu$ m)	$\geq$ 95%	$\geq$ 98%	$\geq$ 98%
ASTM F2299: PFE (Filtration 1 $\mu$ m)	$\geq$ 95% @ 0.1 micron	$\geq$ 98% @ 0.1 micron	$\geq$ 98% @ 0.1 micron
16 CFR Part 1610: Flame Spread (Flammability)	Class 1	Class 1	Class 1

# FOLIA WATER TEST REPORT

## SCOPE OF WORK

Performance Testing of Face Masks to  
ASTM F2100 *Standard Specification for Performance of  
Materials Used in Medical Face Masks*, 2020 Edition

## REPORT NUMBER

104488142CRT-001b

## ISSUE DATE

February 9, 2021

## REVISION DATE:

February 12, 2021

## PAGES

14

## DOCUMENT CONTROL NUMBER

GFT-OP-10i (28-Nov-2018)

© 2021 INTERTEK



## TEST REPORT

Issued February 12, 2021

Intertek Report No. 104488142CRT-001b  
Intertek Project No. G104488142

### CLIENT

FOLIA WATER  
68 34th Street, C642  
Brooklyn, NY 11232  
USA

### TEST STANDARD

ASTM F2100 *Standard Specification for Performance of Materials Used in Medical Face Masks*,  
2020 Edition

### AUTHORIZATION

Quote Number: Qu-01115268-1

### SAMPLE IDENTIFIED BY THE CLIENT AS

Product Type: Face Masks, Disposable  
Brand Name: Folia Water  
Model Numbers: 1024220 A, B, C

### SAMPLE INFORMATION

Date(s) Samples Received: November 17, 2020  
Condition of Samples: Production Run  
Date(s) of Testing: November 23, 2020 Through February 8, 2021

### TEST INFORMATION

ASTM F2101 *Bacterial Filtration Efficiency*  
EN 14683:2019 Annex C *Differential Pressure*  
ASTM F2299 *Sub-Micron Particulate Filtration*  
ASTM F1862 *Resistance to Penetration by Synthetic Blood*  
16 CFR 1610 *Flammability*  
EN 14683:2019 *Microbial Cleanliness (Bioburden)*  
Non-Standard Viral Filtration Efficiency

### STATUS

Test data attached  
Testing in process  
Test data attached  
Test data attached  
Test data attached  
Test data attached  
Test data attached

### TESTING LOCATION

Intertek-Cortland, NY  
Intertek-Guatemala  
Intertek-Guatemala  
Intertek-Cortland, NY  
Intertek-Cortland, NY  
Nelson Labs  
Nelson Labs

This report is for the exclusive use of Intertek's Client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this report. Only the Client is authorized to permit copying or distribution of this report and then only in its entirety. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. The observations and test results in this report are relevant only to the sample tested. This report by itself does not imply that the material, product, or service is or has ever been under an Intertek certification program.

## TEST REPORT

Issued February 12, 2021

Intertek Report No. 104488142CRT-001b

Intertek Project No. G104488142

### SECTION 1

#### CONCLUSION

This test report represents the testing covered by proposal number Qu-01115268-1.

The observations and test results in this report are relevant only to the sample tested. Intertek makes no representations or warranties, express or implied, regarding units that were not tested including, but not limited to, units that may be part of the same lot.

If there are any questions regarding the results contained in this report, or any other services offered by Intertek, please do not hesitate to contact the undersigned.

Please note this Test Report does not represent authorization for the use of any Intertek certification marks.

**Project Owner:** Steven Morey

**Title:** Techincian

**Signature:** 

**Date:** February 12, 2021

**Project Reviewer:** Jason Allen

**Title:** Technical Advisor

**Signature:** 

**Date:** February 12, 2021

#### REPORT REVISIONS

Date / Project #	Project Handler/ Reviewer	Description of Change
2/12/2021	Steven Morey / Jason Allen	Replaced PFE Test Report with Revised Test Report

TEST REPORT

SECTION 2

ASTM F2100-20 TEST DATA

BACTERIAL FILTRATION EFFICIENCY (BFE), ASTM F2101-19

Specimens conditioned for 4-hours at 20.4-22.1°C and 83-86%RH

Test Set-up Information	
Area of Test Specimen (cm <sup>2</sup> )	48.3
Specimen Side Facing Challenge	Inside of Mask
Flow Rate (LPM)	28.3
Averaged + Control Plate Count	2542
Mean Particle Size (µm)	3

Medical Face Mask Barrier Testing					
Plate Count	Mask Specimen				
Stage	1	2	3	4	5
Stage 1	0	0	0	0	0
Stage 2	0	0	0	0	0
Stage 3	0	0	0	0	0
Stage 4	0	2	1	8	0
Stage 5	3	6	3	7	0
Stage 6	1	2	1	1	0
Plate Count Total	4	10	5	16	0
% BFE	99.84	99.61	99.80	99.37	>99.9

TEST EQUIPMENT INFORMATION

Description	Control Number	Calibration Date	Calibration Due
Conditioning Chamber	308-H252	2/26/2020	2/26/2021
Timer	308-H358	1/13/2020	1/13/2021
Pipette	308-H294	2/26/2020	2/26/2021
Analytical Balance	308-S268	12/2/2019	12/2/2020

Date of Testing	11/23/2020
-----------------	------------

## TEST REPORT

### SECTION 4

#### ASTM F2100-20 TEST DATA

#### PARTICULATE FILTRATION EFFICIENCY (PFE), ASTM F2299-17



NUMBER : GUAT20026849-REV1

### Original Sample

### Outside Mask



### Inside Mask



Intertek de Guatemala, S.A.  
46 calle 21-53 zona 12, Expobodegas Petapa 46,  
Ofibodega #10, Guatemala City 01012, Guatemala, PBX: +502 2303 6800

## TEST REPORT



TEST REPORT



NUMBER : GUAT20026849-REV1  
DATE : 29-Jan-2021

APPLICANT : Intertek Testing Services NA Inc.  
3993 US Route 11, Cortland, NY, 13045, U.S.A.  
ATTN : Colin King, Matthew Stevens

THIS IS TO SUPERSEDE REPORT  
NO. GUAT20026849  
DATED 02-Dec-2020

Sample Description : G104488142\_Folia Water\_ Brown  
Buyer : Not Provided  
Vendor : Not Provided  
Agent : Not Provided  
Manufacturer / Factory : Not Provided  
Color : Brown  
Style Number : Not Provided  
PO Number : Not Provided  
Material /Fiber Content : Not Provided  
Standard : Not Provided  
Product end use : Non-medical face mask  
Date Sample Received / Date Test Started : 01 Dec 2020

---

TEST CONDUCTED : AS PER THE REQUEST OF THE APPLICANT. FOR FURTHER DETAILS PLEASE REFER TO ENCLOSED PAGE(S)

---

### CONCLUSION :

α Determination of the Initial Efficiency of Materials Used in Medical Face Masks to Penetration by Particulates Using Latex Spheres D

---

NOTE: P = Meet Buyer's Requirement F = Below Buyer's Requirement  
\* = See Remark C = Conform Label  
N/A = Not Applicable D = Data

This report was issued in order to correct the pictures.

AUTHORIZED BY  
FOR Intertek de Guatemala, S.A. [Guatemala]

RUDY SEMRAU  
GENERAL MANAGER

REPORTES GUATEMALA

1. Any copying or replication of this report to or for any other person or entity, or use of our name or trademark, is permitted only with our prior written permission. 2. Our report includes all of the tests requested by you and the results thereof based upon the information that you provided to us. You have 60 days from the date of issuance of this report to notify us of any material error or omission caused by our negligence, provided, however, that such notice shall be in writing and shall specifically address the issue you wish to raise. A failure to raise such issue within the prescribed time shall constitute your unqualified acceptance of the completeness of this report, the tests conducted and the correctness of the



## Latex Particle Challenge Final Report

---

Test Article: 102420A - 7 samples  
102420B - 7 samples  
102420E - 7 samples  
Study Number: 1397030-S01  
Study Received Date: 10 Mar 2021  
Testing Facility: Nelson Laboratories, LLC  
6280 S. Redwood Rd.  
Salt Lake City, UT 84123 U.S.A.  
Test Procedure(s): Standard Test Protocol (STP) Number: STP0005 Rev 08  
Deviation(s): None

**Summary:** This procedure was performed to evaluate the non-viable particle filtration efficiency (PFE) of the test article. Monodispersed polystyrene latex spheres (PSL) were nebulized (atomized), dried, and passed through the test article. The particles that passed through the test article were enumerated using a laser particle counter.

A one-minute count was performed, with the test article in the system. A one-minute control count was performed, without a test article in the system, before and after each test article. Control counts were performed to determine the average number of particles delivered to the test article. The filtration efficiency was calculated using the number of particles penetrating the test article compared to the average of the control values. During testing and controls, the air flow rate is maintained at 1 cubic foot per minute (CFM)  $\pm$  5%.

The procedure employed the basic particle filtration method described in ASTM F2299, with some exceptions; notably the procedure incorporated a non-neutralized challenge. In real use, particles carry a charge, thus this challenge represents a more natural state. The non-neutralized aerosol is also specified in the FDA guidance document on surgical face masks. All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Test Side: Inside  
Area Tested: 91.5 cm<sup>2</sup>  
Particle Size: 0.1  $\mu$ m  
Laboratory Conditions: 16 Mar 2021: 21.2°C, 22% relative humidity (RH) at 2204;  
21.1°C, 22% RH at 2252; 20.9°C, 22% RH at 2308  
21 Mar 2021: 21.2°C, 22% RH at 1401; 21.0°C, 22% RH at 1446



---

Cameron Brierley electronically approved  
Study Director

Cameron Brierley

---

25 Mar 2021 16:48 (+00:00)  
Study Completion Date and Time

**Results:**
Test Article: 102420A

Test Article Number	Test Article Counts	Average Control Counts	Filtration Efficiency (%)
1	43	13,392	99.68
2	34	13,340	99.75
3	38	13,131	99.71
4	38	12,764	99.70
5	38	12,922	99.71
6	32	12,504	99.74
7	34	11,547	99.71

 Average Filtration Efficiency: 99.71%  
 Standard Deviation: 0.024

Test Article: 102420B

Test Article Number	Test Article Counts	Average Control Counts	Filtration Efficiency (%)
1	37	11,995	99.69
2	30	12,466	99.76
3	41	12,135	99.66
4	45	11,434	99.61
5	28	11,785	99.76
6	45	11,365	99.60
7	44	11,019	99.60

 Average Filtration Efficiency: 99.67%  
 Standard Deviation: 0.071

Test Article: 102420E

Test Article Number	Test Article Counts	Average Control Counts	Filtration Efficiency (%)
1	52	12,331	99.58
2	37	13,258	99.72
3	40	13,731	99.71
4	62	14,017	99.56
5	44	14,341	99.69
6	64	14,715	99.57
7	22	12,463	99.82

 Average Filtration Efficiency: 99.66%  
 Standard Deviation: 0.100

TEST REPORT

SECTION 5

ASTM F2100-20 TEST DATA

RESISTANCE TO PENETRATION BY SYNTHETIC BLOOD, ASTM F1862-17

Specimens conditioned for 4-hours at 21-22°C and 83-86%RH

Specimens tested at 21-23°C and 55-62% RH

Medical Face Mask Barrier Testing				
Specimen	Pressure	Test Volume (mL)	Visible Penetration of Blood or Wetness	Pass/Fail
1	160 mmHg	2	None	Pass
2	160 mmHg	2	None	Pass
3	160 mmHg	2	None	Pass
4	160 mmHg	2	None	Pass
5	160 mmHg	2	None	Pass
6	160 mmHg	2	None	Pass
7	160 mmHg	2	None	Pass
8	160 mmHg	2	None	Pass
9	160 mmHg	2	None	Pass
10	160 mmHg	2	None	Pass
11	160 mmHg	2	None	Pass
12	160 mmHg	2	None	Pass
13	160 mmHg	2	None	Pass
14	160 mmHg	2	None	Pass
15	160 mmHg	2	None	Pass
16	160 mmHg	2	None	Pass
17	160 mmHg	2	None	Pass
18	160 mmHg	2	None	Pass
19	160 mmHg	2	None	Pass
20	160 mmHg	2	None	Pass
21	160 mmHg	2	None	Pass
22	160 mmHg	2	None	Pass
23	160 mmHg	2	None	Pass
24	160 mmHg	2	None	Pass
25	160 mmHg	2	None	Pass
26	160 mmHg	2	None	Pass
27	160 mmHg	2	None	Pass
28	160 mmHg	2	None	Pass
29	160 mmHg	2	None	Pass
30	160 mmHg	2	None	Pass
31	160 mmHg	2	None	Pass
32	160 mmHg	2	None	Pass

## TEST REPORT

## TEST EQUIPMENT INFORMATION

Description	Control Number	Calibration Date	Calibration Due
Conditioning Chamber	308-H252	2/26/2020	2/26/2021
Automated Dispenser	308-H386	VBU	VBU
Ambient Conditions Monitor	308-G183	4/21/2020	4/21/2021
Timer for Dispenser Verification	308-N1257	3/12/2020	3/12/2021

Date of Testing	12/7/2020
-----------------	-----------

TEST REPORT

SECTION 6

16 CFR 1610 TEST DATA

FLAMMABILITY OF CLOTHING TEXTILES

Surface type: Plain, Single Layer  
Tested side: Face

Preliminary Test - Original State	
Length Direction	Burn Time (s)
Up	IBE
Down	9.85
Width Direction	Burn Time (s)
Up	9.86
Down	10.23

Final Test - Original State Width Up Direction	
Specimen	Burn Time (s)
1	8.68
2	11.12
3	12.68
4	8.98
5	9.86

**Classification:** Class 1, Normal Flammability

Note: Sample is one-time use item, flammability testing performed in original state only

Test Result Codes: Plain Surface Fabrics	
DNI	Did not ignite (no time)
IBE	Ignited, but extinguished (no time)

TEST EQUIPMENT INFORMATION

Description	Control Number	Calibration Date	Calibration Due
Circulating Oven	308-H223	2/26/2020	2/26/2021
Flame Chamber	US20041501	VBU	VBU

**Date of Testing** 12/4/2020

TEST REPORT

SECTION 6

EN 14683:2019 TEST DATA

MICROBIAL CLEANLINESS (BIOBURDEN)



Sponsor:  
Colin King  
Intertek Etl Semko Division  
3933 Us Rte. 11  
Cortland, NY 13045

Microbial Cleanliness (Bioburden) of Medical Masks Final Report

Test Article: G104488142\_FW\_Brown\_MASK  
Purchase Order: USA20-0000246160  
Study Number: 1370695-S01  
Study Received Date: 09 Dec 2020  
Testing Facility: Nelson Laboratories, LLC  
6280 S. Redwood Rd.  
Salt Lake City, UT 84123 U.S.A.  
Test Procedure(s): Standard Test Protocol (STP) Number: STP0036 Rev 15  
Customer Specification Sheet (CSS) Number: 202002096 Rev 01  
Deviation(s): None

Summary: The testing was conducted in accordance with EN 14683:2019 and ANSI/AAMVISO 11737-1:2018. When bioburden results are calculated using a software program, manual calculations may differ slightly due to rounding. The counts determined on products are colony forming units and may not always reflect individual microorganisms. The sponsor performs any statistical analysis and determines the acceptable limits. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Results:

Unit Number	Weight (g)	Aerobic	Fungal	Total Bioburden (CFU/mask)	Total Bioburden (CFU/g)
1	3.7	<3	<3	<6.1	<1.7
2	3.6	<3	<3	<6.0	<1.7
3	3.8	<3	<3	<6.2	<1.6
4	3.6	<3	<3	<6.0	<1.7
5	3.6	<3	<3	<6.0	<1.7
Recovery Efficiency	UTD*				

< = No Organisms Detected

UTD = Unable to Determine

Note: The results are reported as colony forming units (CFU) per mask.

\* UTD due to zero count on the first rinse. An alternative method or inoculated product recovery efficiency is recommended.



Gabrielle Waldron electronically approved  
Study Director

Gabrielle Waldron

18 Dec 2020 16:54 (+00:00)  
Study Completion Date and Time

## TEST REPORT



Study Number 1370895-S01  
Microbial Cleanliness (Bioburden) of Medical Masks Final Report

### Method Suitability:

Organism	Percentage
<i>Bacillus atrophaeus</i>	0%

**Test Method Acceptance Criteria:** If applicable, anaerobic controls are acceptable for the bioburden test results. The number of masks to be tested shall be a minimum of 5. The bioburden of the medical mask shall be  $\leq 30$  CFU/g tested.

### Procedure:

Positive Controls/Monitors: *Bacillus atrophaeus*  
Extract Fluid: Peptone Tween<sup>®</sup>  
Extract Fluid Volume: ~300 mL  
Extract Method: Orbital Shaking  
Plating Method: Membrane Filtration  
Agar Medium: Tryptic Soy Agar  
Potato Dextrose Agar  
Recovery Efficiency: Exhaustive Rinse Method  
Aerobic Bacteria: Plates were incubated 3-7 days at 30-35°C, then enumerated.  
Fungal: Plates were incubated 5-7 days at 20-25°C, then enumerated.

## TEST REPORT

### SECTION 7

#### NON-STANDARD VIRAL FILTRATION EFFICIENCY



Sponsor:  
Krissie Brown  
Intertek Etl Semko Division  
3933 US Route 11  
Cortland, NY 13045

### Viral Filtration Efficiency (VFE) Final Report

Test Article: Folia Water G104488142 VFE  
Purchase Order: USA20-0000246336  
Study Number: 1377843-S01  
Study Received Date: 06 Jan 2021  
Testing Facility: Nelson Laboratories, LLC  
6280 S. Redwood Rd.  
Salt Lake City, UT 84123 U.S.A.  
Test Procedure(s): Standard Test Protocol (STP) Number: STP0007 Rev 16  
Deviation(s): None

**Summary:** The VFE test is performed to determine the filtration efficiency of test articles by comparing the viral control counts upstream of the test article to the counts downstream. A suspension of bacteriophage ΦX174 was aerosolized using a nebulizer and delivered to the test article at a constant flow rate and fixed air pressure. The challenge delivery was maintained at  $1.1 - 3.3 \times 10^3$  plaque forming units (PFU) with a mean particle size (MPS) of  $3.0 \mu\text{m} \pm 0.3 \mu\text{m}$ . The aerosol droplets were drawn through a six-stage, viable particle, Andersen sampler for collection. The VFE test procedure was adapted from ASTM F2101.

All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Test Side: Inside  
Test Area:  $\sim 40 \text{ cm}^2$   
VFE Flow Rate: 28.3 Liters per minute (L/min)  
Conditioning Parameters:  $85 \pm 5\%$  relative humidity (RH) and  $21 \pm 5^\circ\text{C}$  for a minimum of 4 hours  
Positive Control Average:  $1.1 \times 10^3$  PFU  
Negative Monitor Count:  $< 1$  PFU  
MPS:  $2.9 \mu\text{m}$



Janelle Bentz electronically approved for  
Study Director

Mikell Goldsbery

08 Feb 2021 19:50 (+00:00)

Study Completion Date and Time



## TEST REPORT



Study Number 1377843-S01  
Viral Filtration Efficiency (VFE) Final Report

### Results:

Test Article Number	Percent VFE (%)
1	99.6
2	99.3
3	99.3
4	99.5
5	99.7

The filtration efficiency percentages were calculated using the following equation:

$$\% VFE = \frac{C - T}{C} \times 100$$

C = Positive control average

T = Plate count total recovered downstream of the test article

Note: The plate count total is available upon request